

DRRR GmbH · Bodmanstraße 4 · D-87435 Kempten

Mrs Alice Palma
Agrumaria Corleone spa
Via S. Corleone 12
90124 Palermo
Italy

DRRR-Proficiency Testing
RVEP 180712 fruit juice or fruit juice concentrate 3

Dear participant

Thank you very much for participating in our proficiency testing.

Your lab code number is: 5

We offer the test material of this proficiency testing as a reference material. If you are interested, we would be very pleased to submit you our offer.

We hope, that we will meet your laboratory in the next your proficiency testing round again. Registrations for our program are possible any time using our online system (ODIN) or the registration forms on our homepage: www.DRRR.de

For questions do not hesitate to contact us.

Yours sincerely

Benjamin Bartl

date: 05.12.2018

your contact persons:
Benjamin Bartl
Norman Walter

fon: +49 (0)831/960 878-0
fax: +49 (0)831/960 878-99
e-mail: info@DRRR.de

Deutsches Referenzbüro für Ringversuche und
Referenzmaterialien GmbH
Bodmanstraße 4 D-87435 Kempten
fon: +49 (0)8 31/960 878-0
fax: +49 (0)8 31/960 878-99
e-Mail: info@DRRR.de
website: www.DRRR.de

accounting details:
VAT-no DE254613132 •TAX-no 127/124/32207
Raiffeisenbank im Allgäuer Land
bank code 733 692 64 *account no 102350
IBAN DE 94733692640000102350
BIC-Code: GENO DEF1DTA

place of business: Kempten
commercial registry: HRB 9496
local court: Kempten
managing director:
Dr. rer. nat. Ulrich Leist
Thorsten Helbig M.Eng.

CERTIFICATE FOR PARTICIPATION

DRRR Proficiency Testing

The participant

Agrumaria Corleone spa
Via S.Corleone 12
90124 Palermo
Italy

Customer number: 12088

lab code number: 5

gets for the above mentioned quantitative proficiency testing the following valuation:

great success

To the above mentioned proficiency testing the corresponding report contains the statistical evaluations.

Kempton, 04.12.2018



Benjamin Bartl

Valuation:

If the mean of z'-scores of all evaluated results is smaller than 2 or equal and single z'-scores are smaller than 2 or equal you will get the valuation: „great success“

If the mean of z'-scores of all evaluated results is smaller than 2 or equal and single z'-scores are smaller than 2 or equal you will get the valuation: „success“

If the mean of z'-scores of all evaluated results is higher than 2 you will get the valuation: „participated“

Summary of your quantitative lab performance according to the reference methods

For your company

Agrumaria Corleone spa

SUMMARY

customer number: 12088
lab code number: 5
number of z'-scores: 10
number of z'-scores > 2: 0

Valuation:

If the mean of z'-scores of a parameter is smaller than 2 or equal you will get the valuation:
"You performed very well. Keep your performance."

If the mean of z'-scores of a parameter is smaller than 2 or equal but at least one single z'-score is higher than 2 you will get the valuation:
"You performed well. In particular cases you showed potential for improvement."

If the mean z'-score of a parameter is higher than 2 you will get the valuation:
"You showed potential for improvement of your performance."

parameter	number of z'-scores:	number of z'-scores > 2:	mean of z'-scores:	Valuation:
brix value	2	0	0,30	You performed very well. Keep your performance.
pH-value	2	0	1,18	You performed very well. Keep your performance.
titratable acidity (pH 8,1)	1	0	0,40	You performed very well. Keep your performance.
ash	-	-	-	-
potassium	2	0	0,04	You performed very well. Keep your performance.
magnesium	1	0	0,42	You performed very well. Keep your performance.
phosphorus	-	-	-	-
sodium	2	0	1,50	You performed very well. Keep your performance.
copper	-	-	-	-

In this summary all your z'-score are selected for the evaluation. The z'-scores refer to the reference method. For routine reports we do not generate a performance summary for all used methods. In particular cases we can generate you a special performance summary with your used method.